



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/633,299	08/01/2003	Assaf Govari	BIO-5023	3322

27777 7590 02/12/2007
PHILIP S. JOHNSON
JOHNSON & JOHNSON
ONE JOHNSON & JOHNSON PLAZA
NEW BRUNSWICK, NJ 08933-7003

EXAMINER

KISH, JAMES M

ART UNIT

PAPER NUMBER

3737

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	02/12/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary

Application No.

10/633,299

Applicant(s)

GOVARI, ASSAF

Examiner

James Kish

Art Unit

3737

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 23-37 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 23-37 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 01 August 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. ____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>3/1/05, 6/8/06</u> | 6) <input type="checkbox"/> Other: ____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

1. Claims 23-24 and 26-37 rejected under 35 U.S.C. 102(b) as being anticipated by Osadchy et al. (US Patent No. 6,266,551). Osadchy discloses a probe for insertion into the body of a subject and including an electronic microcircuit, which stores information relating to calibration of the probe. The calibration information includes all or any combination of the following: data relating to deviation of the coils from orthogonality, data relating to the respective gains of the coils and/or data relating to the relative displacement of the distal tip from the coils (column 2, line 64 through column 3, line 4). This calibration information includes data relating to signal non-linearities (column 3, lines 19-22), data calculated by differences between signals generated by a first and second master coil (column 14, lines 40-54) and the calibration data may also relate to a proportionality to a directional component of the magnetic fields (column 9, lines 4-13). Regarding memory, see column 5, lines 21-32 and column 7, lines 21-39. Alternatively or additionally, at least some of the calibration data are determined by applying spatially variable magnetic fields to the probe (column 9, lines 6-8). For further description of the Osadchy reference, see column 2, lines 18-63 and column 4, lines 7-27.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claim 25 is rejected under 35 U.S.C. 103(a) as being unpatentable over Osadchy in view of Strommer et al. (US Patent No. 6,233,476). Osadchy discloses a probe for insertion into the body of a subject and including an electronic microcircuit, which stores information relating to calibration of the probe, as described in the rejection of claims 23-24 and 26-37. Osadchy does not explicitly state that the position and orientation sensor can be utilized in a capsule, however, it is disclosed that the catheter is a wireless catheter which is not physically connected to the signal processing and/or computing apparatus. Rather, a transmitter/receiver is attached to a proximal end of the catheter. One benefit of this type of configuration is that the catheter, which is inserted into the heart, can easily be made electrically floating. Strommer teaches a method and system for initiating and calibrating the location and orientation of an objects and a detector system within a scanning volume. It is noted that the housing for the positioning object can be shaped like a capsule (column 3, lines 44-52). It would have been obvious to one having ordinary skill in the art at the time the invention was made to broaden the scope of Osadchy by placing the positioning device into a capsule as taught by Strommer because even within the catheter the system is electrically floating

Art Unit: 3737

and isolated, thereby making the capsule configuration an obvious variant of that of the catheter.

Conclusion

Other related prior art:

Schneider	6,073,043
Hossack et al.	6,338,716
Ben-Haim et al.	6,690,963
Frisch et al.	6,904,308
Mizuno	6,939,292


Any inquiry concerning this communication or earlier communications from the examiner should be directed to James Kish whose telephone number is 571-272-5554. The examiner can normally be reached on 8:30 - 5:00 ~ Mon. - Fri..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian Casler can be reached on 571-272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3737

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JMK


EILEEN M. HERDER
SUPERVISOR, PATENT EXAMINER